



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2022



PC-1024

<b>Lab Name</b>	<b>REDCLIFE LIFETECH PVT LTD</b>	<b>Lab No</b>	<b>15536</b>
<b>Constituent Group</b>	<b>HbA1c</b>	<b>Date of Result Entered :</b>	<b>13/07/2022</b>
<b>PT item</b>	<b>Lyophilized Whole blood</b>	<b>Date of Report Published :</b>	<b>05/08/2022</b>

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	HbA1c	Enzymatic	Abbott	18	4.09	6.75	0.28	3.9 %	-0.69	0.13

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

**REDCLIFE LIFETECH PVT LTD**  
**2ND FLOOR, 47, PANCHWATI, SAHELI MARG**  
**UDAIPUR**  
**RAJASTHAN313001**

**Coordinator Contact Details:**  
**Email: clinqc@cmcvellore.ac.in**  
**Contact Number: 0416-2283102**

*Pamela Christudoss*  
**Dr. Pamela Christudoss**  
**CMC EQAS Coordinator**  
**Christian Medical College, Vellore**

**Homogeneity and Stability of the sample is passed.**  
**Data in CMC EQAS reports is confidential**  
**CMC EQAS does not sub contract any components**  
**\*\*\*\*\* End of Report \*\*\*\*\***