



## **Abbott (Diagnostics Division)**

405, KIRTI SHIKHAR BUILDING,  
JANAKPURI DISTT. CENTRE,  
NEW DELHI - 110 058  
TEL : +91 11 25619194 FAX: +91 11 25613392

Dated-Dec 17<sup>th</sup> , 2021

### **Certificate Of Calibration**

This is to certify that the below mentioned instrument installed at Redcliffe lifetech Pvt Ltd, Udaipur calibrated by us on 17th Dec, 2021. The next calibration is due on or before 16<sup>th</sup> Dec, 2022. The values are falling within the specified range.

Instrument Model- C4000  
Serial No. -C403079  
Software Version - 9.45

  
For Abbott Healthcare Pvt Ltd

**Prashant Mishra**  
**Technical Service Specialist**  
**Abbott Diagnostic Division**

# QC Results List Report

Unreleased

Operator ID:

System serial number: 85362

C / P	Module	SID	Control name	Assay	Result	Flags	Code	Date / Time
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	UIBC-L	53 ug/dL			14.08.2022 09:31
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	Fe-PI	255 ug/dL			14.08.2022 09:31
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	UA	4.6 mg/dL			14.08.2022 09:31
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	ALT	27 U/L			14.08.2022 09:31
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	GGT	65 U/L			14.08.2022 09:30
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	UHDL	76 mg/dL			14.08.2022 09:30
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	BiID	0.42 mg/dL			14.08.2022 09:29
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	AST	36 U/L			14.08.2022 09:29
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	Urea	15.3 mg/dL			14.08.2022 09:29
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	Chol	254 mg/dL			14.08.2022 09:28
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	CreaC	2.48 mg/dL			14.08.2022 09:28
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	AlkP	94 U/L			14.08.2022 09:28
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	Phos	3.4 mg/dL			14.08.2022 09:27
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	BillIT	1.03 mg/dL			14.08.2022 09:26
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	GluC	75 mg/dL			14.08.2022 09:24
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	TP	6.36 g/dL			14.08.2022 09:24
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	Trig	182 mg/dL			14.08.2022 09:23
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	CaC	9.3 mg/dL			14.08.2022 09:22
W968/1	1	BIO-CHELevel 1	BIO-CHE - Level 1	AlbG	3.90 g/dL			14.08.2022 09:22

*Adelish*  
14/08/22

# QC Results List Report

Unreleased

Operator ID:

System serial number: 85362

C / P	Module	SID	Control name	Assay	Result	Flags	Code	Date / Time
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	UIBC-L	121 ug/dL			14.08.2022 09:34
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	Fe-PI	70 ug/dL			14.08.2022 09:34
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	UA	9.5 mg/dL			14.08.2022 09:34
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	ALT	98 U/L			14.08.2022 09:34
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	GGT	159 U/L			14.08.2022 09:33
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	UHDL	24 mg/dL			14.08.2022 09:33
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	BiID	1.76 mg/dL			14.08.2022 09:32
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	AST	172 U/L			14.08.2022 09:32
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	Urea	46.0 mg/dL			14.08.2022 09:32
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	Chol	103 mg/dL			14.08.2022 09:31
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	CreaC	5.95 mg/dL			14.08.2022 09:31
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	AlkP	384 U/L			14.08.2022 09:31
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	Phos	7.1 mg/dL			14.08.2022 09:30
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	BiIT	4.24 mg/dL			14.08.2022 09:29
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	GluC	265 mg/dL			14.08.2022 09:27
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	TP	3.97 g/dL			14.08.2022 09:27
68/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	Trig	90 mg/dL			14.08.2022 09:26
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	CaC	11.8 mg/dL			14.08.2022 09:25
W968/2	1	BIO-CHELevel 2	BIO-CHE - Level 2	AlbG	2.51 g/dL			14.08.2022 09:25

*Handwritten signature*  
14/08/22

# QC Results List Report

Operator ID:

Unreleased

System serial number: 85362

C / P	Module	SID	Control name	Assay	Result	Flags	Code	Date / Time
R462/2	5	BIO-HbA1cLevel 2	BIO-HbA1c - Level 2	%A1cH	9.0 %			14.08.2022 09:36
R462/1	5	BIO-HbA1cLevel 1	BIO-HbA1c - Level 1	%A1cH	4.7 %			14.08.2022 09:35

*Abhishek*  
14/08/22