



Ref No. TS/6618/SBP

Date : 8<sup>th</sup> March '2022

### Calibration report

- 1.Instrument- Easylyte Plus Sr No.C66417CNKC
- 2.Year of manufacturing : 2022
- 3.Date of Calibration - 5th march '2022

Test Voltages / Test Temperature -

Test Voltage	Actual Voltage	OK/not OK
Logic 5 V	5.06V	OK
Line voltage	230V	OK
Logic 12V	12V	OK
Calibration Value (Na)	59.45	OK
Calibration Value (K)	59.71	OK
Calibration Value (Li)	49.04	OK



Erba Corporate Services Ltd . Registered in England and Wales . Company No: 11100097  
Registered Office: c/o Erba Molecular, Unit 4 Bartholemews, Ely, Cambridgeshire, CB74EA, UK



Calibration Value (Li)	49.04	OK
Electrode Values(Na)	131	OK
Electrode Value(K)	85	OK
Electrode Value(Cl)	57	OK
Pump calibration		OK
Fluid Test		OK

Instrument checked with Medica EasyQC Controls

Thanking you and assuring of our prompt attention at all times.


Yours faithfully,  
For **TRANSASIA BIO-MEDICALS LTD.**



Product manager,

Electrolyte analyzer

Erba Corporate Service Ltd.  
GW1 Great West House  
Great West Road, London TW8 9DF



16/08/22

CALIBRATION

CAL VALUES

Na 57.37 K 58.73 Cl 46.08  
AUG-16-22; 09:57

ANALYSIS

PATIENT'S NAME

OC-I

SAMPLE 001

Na 139.8 K 3.81 Cl 103.8  
BLOOD Na,K,Cl mmol/L (ISE)  
AUG-16-22; 09:58

ANALYSIS

PATIENT'S NAME

OC-II

SAMPLE 002

Na 160.1 K 6.02 Cl 123.5  
Na HI K HI Cl HI  
BLOOD Na,K,Cl mmol/L (ISE)  
AUG-16-22; 10:00

ANALYSIS

PATIENT'S NAME

OC-III

SAMPLE 003

Na 122.9 K 2.87 Cl 84.5  
Na LO K LO Cl LO  
BLOOD Na,K,Cl mmol/L (ISE)  
AUG-16-22; 10:03

*16/08/22*