


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 220808A/N-110	Page	: 1 of 1
CALIBRATED ON.	: 09 / 08 / 2022	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 08 / 08 / 2023	C.P. No.	: ACS/CP/10
SRF No. & DATE	: 220808A/N Dt. 08.08.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342722000000223F Dt. 10.08.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 2 nd Floor 47, Panchwati, Saheli Marg, Udaipur- 313001		
ITEM CALIBRATED	: Centrifuge	Make	: Remi
		Range	: (0 to 6000) RPM
		L. Count	: ID No. : RCL/CP/Cent-01
		Location	: Clinical Pathology
		S. No.	: ZHHN-33091
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C} R.H. {40% to 60%}		
STANDARD INSTRUMENTS USED :			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.
Tachometer	ACS/DTM/01/ TM5010EK-039	Godrej	M-210914-16-1
			DUE DATE
			14.09.2022
RESULTS	: Mechanical Discipline		
* U. C. INSTT.	Standard Measured	Error	
READS (RPM)	(RPM)	(RPM)	
1000	1001.60	- 1.60	
2000	2002.20	- 2.20	
3000	3002.80	- 2.80	
4000	4004.50	- 4.50	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.

***U.C.** : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

