



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com **CC-3427**



Calibration Certificate

CERTIFICATE NO.	: 220808A/N-112	Page	: 1 of 1	
CALIBRATED ON.	: 09 / 08 / 2022	Ref. Std.	: IS:12508:1988	
CALIBRATION DUE.	: 08 / 08 / 2023	C.P. No.	: ACS/CP/10	
SRF No. & DATE	: 220808A/N Dt. 08.08.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000000225F Dt. 10.08.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 2 nd Floor 47, Panchwati, Saheli Marg, Udaipur- 313001			
ITEM CALIBRATED	: Centrifuge			
	Make : Remi	Range	: (0 to 6000) RPM	
	L. Count :	ID No.	: RCL/HEM/Cent-03	
	Location : Haematology	S. No.	: ZHKN-42555	
ENVIRONMENT CONDITION:	Temp. { 25°C + 3°C } R.H. { 40% to 60% }			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ TM5010EK-039	Godrej	M-210914-16-1	14.09.2022
RESULTS : Mechanical Discipline				
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)		
1000	1001.50	- 1.50		
2000	2002.70	- 2.70		
3000	3003.90	- 3.90		
4000	4004.30	- 4.30		
(average of readings)				

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor $k=2$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

