

Date -3rd August 2022

Calibration Certificate

This is to certify that Fully Automatic Biochemistry Analyzer “Sphera” with serial no. 201457002 at SRL Limited, Panjim-Goa has been successfully inspected & calibrated.

The following are the calibration details are as follows –

| | |
|---------------------------------|------------------------|
| Installation Date | 06/08/2021 |
| Calibration Date | 03/08/2022 |
| Calibration due date | 03/08/2023 |
| Electrical and Functional check | Found OK |
| Hardware check | Found OK |
| Mechanical calibration Check | Found OK |
| Wash Function check | Found OK |
| Sample Arm movement check | Found OK |
| Optical Reading test | Found OK |
| Parameters check | Controls run, found ok |

| Wavelength | mvolt | Efficiency |
|------------|--------|------------|
| 340 | 1023.5 | 97% |
| 405 | 1411.5 | 99% |
| 505 | 1343.2 | 100% |
| 546 | 1452.5 | 98% |
| 578 | 1368 | 96% |
| 600 | 1494.9 | 100% |
| 650 | 1487.9 | 99% |
| 700 | 1376.4 | 96% |

For Accurex Biomedical Pvt LTD.
Regional Service Engineer- Mr. Prashant More

212 Udyog Mandir No.1,7/C,Bhagojee Keer Marg,Mahim,Mumbai-400016, Ph.: (022) 6744 6744



