



EURO CALIBRATION TECHNIQUE

No. 387, IInd Floor, TIC Road, Varadharajapuram,
Ambattur, Chennai-600 053. Tamil Nadu

Tel: 044-48521288 Mobile: 9486195184, 9176032184

E-mail: eurocalibration@gmail.com web: www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2022/0642- 02

ULR Number : CC2519 22 0 00006087F

Customer Name and Address

M/s. Government Vellore Medical College & Hospital,
Room No: 16, OP Block, Addukamparai,
Vellore - 632 011.

Date of Issue : 9-Aug-2022

Pages : 1 of 1

SRF Number : 2022/0642

Reciept Date : 6-Aug-2022

Date of Calibration : 6-Aug-2022

Suggested calibration Due Date : 5-Aug-2023

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Icelined Refrigerator

Make : Vestfrost

Model : VLS 354A AC

S.No : 20210716170

ID No. : GVMCH/ICTC/016

Range : 2 to 8 °C

Resolution : 1 °C

Accuracy : As Per Manual

Location : --

Environmental Conditions

Temperature : 25.6 °C

Relative Humidity : 54 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor	1702	ECT 2021/0826-03	05.11.2022
2	Digital RTD Calibrator	11110002959	CT21-438-1	15.12.2022

Calibration Results

Parameter - Thermal / Temperature

S.No	Test Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Uncertainty ±°C
		1	2	3		
1	5	5.4	5.5	5.5	5.47	0.65