



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2022



PC-1024

Lab Name: GOVT PRIMARY HEALTH CENTER

Lab No: 12905

Constituent Group: Chemistry III

Date of Result Entered: 19/03/2022

PT Item: Lyophilized Serum

Date of Report Published: 06/04/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	CV	SD	Your Value	SDI	U		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1766	151.83	7.28	11.06	166 mg/dL	1.28	0.53		
2	CREATININE III	Jafes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1708	1.20	11.54	0.14	1.21 mg/dL	0.07	0.01		
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1655	115.16	6.70	7.71	109 mg/dL	-0.80	0.38		
SDI Range Interpretation												
				Within -1.00 to +1.00							Excellent.	
				Within ±1.01 to ±2.00							Good.	
				Within ±2.01 to ±2.99							Accept with caution. Warning Signal.	
				Beyond ±3.0							Unacceptable performance. Action Signal.	

LAB ADDRESS :
GOVT PRIMARY HEALTH CENTER
THALAVAI PATTINAM
TAMILNADU


Coordinator Contact Details:
Email: cljngc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - APRIL 2022



Lab Name: GOVT PRIMARY HEALTH CENTER
 Constituent Group: Chemistry III
 PT Item: Lyophilized Serum

Lab No: 12905
 Date of Result Entered: 20/04/2022
 Date of Report Published: 05/05/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1781	263.81	7.17	18.93	252.9 mg/dL	-0.58	0.90
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1758	4.67	9.43	0.44	4.48 mg/dL	-0.43	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1741	115.66	6.19	7.16	115.4 mg/dL	-0.04	0.34
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 GOVT PRIMARY HEALTH CENTER
 THALAVAI PATTINAM
 TAMILNADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Paula Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MAY 2022



PC-1024

Lab Name: GOVT PRIMARY HEALTH CENTER

Lab No: 12905

Consultant Group: Chemistry III

Date of Result Entered: 20/05/2022

PT Item: Lyophilized Serum

Date of Report Published: 08/06/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	146.47	6.38	9.35	150 mg/dL	0.38	0.44
2	CREATININE III	Jafes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1771	8.11	8.92	0.72	8.5 mg/dL	0.54	0.03
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1730	119.85	6.17	7.39	126 mg/dL	0.83	0.36
SDI Range Interpretation										
		Within -1.00 to +1.00	Excellent							
		Within ±1.01 to ±2.00	Good.							
		Within ±2.01 to ±2.99	Accept with caution. Warning Signal.							
		Beyond ±3.0	Unacceptable performance. Action Signal.							

LAB ADDRESS :
 GOVT PRIMARY HEALTH CENTER
 THALAVAI PATINAM
 TAMILNADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2022



PC-1024

Lab Name: GOVT PRIMARY HEALTH CENTER

Lab No: 12905

Constituent Group: Chemistry III

Date of Result Entered: 18/06/2022

PT Item: Lyophilized Serum

Date of Report Published: 07/07/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1828	87.43	7.74	6.77	86 mg/dL	-0.21	0.32
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1769	1.99	8.26	0.16	2.1 mg/dL	0.67	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1737	118.45	6.70	7.93	127 mg/dL	1.08	0.38
SDI Range Interpretation										
				Within -1.00 to +1.00	Excellent.					
				Within ±1.01 to ±2.00	Good.					
				Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
				Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS :
GOVT PRIMARY HEALTH CENTER
THALAVAI PATINAM
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JULY 2022



PC-1024

Lab Name GOVT PRIMARY HEALTH CENTER

Lab No

12905

Constituent Group Chemistry III

Date of Result Entered :

20107/2022

PT item Lyophilized Serum

Date of Report Published :

05/08/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1864	295.51	7.14	21.09	307 mg/dL	0.54	0.98
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1839	3.40	9.21	0.31	3.38 mg/dL	-0.06	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1803	117.79	6.19	7.29	120 mg/dL	0.30	0.34
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 GOVT PRIMARY HEALTH CENTER
 THALAVAI PATTINAM
 TAMILNADU

Coordinator Contact Details:
 Email: cllngc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudas
 Dr. Pamela Christudas
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential