



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2022



PC-1024

Lab Name	GOVT PATCHUR COMMUNITY HEALTH CENTRE	Lab No	10642
Constituent Group	Chemistry III	Date of Result Entered :	09/04/2022
PT item	Lyophilized Serum	Date of Report Published :	05/05/2022

S.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1781	263.81	7.17	18.93	246.55 mg/dL	-0.91	0.90
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1758	4.67	9.43	0.44	5.58 mg/dL	2.06	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1741	115.66	6.19	7.16	133.49 mg/dL	2.49	0.34

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

GOVT PATCHUR COMMUNITY HEALTH CENTRE
 CLINICAL LABORATORY ,GOVT PATCHUR COMMUNITY HEALTH CENTER TIRUPATTUR DT-635853.
 PATCHUR
 TAMILNADU635854

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

[Signature]
 Medical Officer,
 Community Health Centre,
 Pachur-635 854.
 Natrampalli Block,
 Tirupattur District.