

## VIEW LAB MONTHLY SUMMARY

Lab Name GOVT PATCHUR COMMUNITY HEALTH CENTRE  
 Lab No 10642  
 Month March  
 Year 2022  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 15/03/2022

Date of Report Published : 06/04/2022

Sl No	Analyte	Method - Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1766	151.83	7.28	11.06	162.36 mg/dL	0.95	0.53
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1708	1.20	11.54	0.14	1.37 mg/dL	1.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1655	115.16	6.70	7.71	134.73 mg/dL	2.54	0.38

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Monthly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*M. S. S. S. S.*  
 Medical Officer,  
 Community Health Centre,  
 Pachur-635 854,  
 Natrampalli Block,  
 Tirupattur District.