



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - FEBRUARY 2022



PC-1024

Lab Name	GOVT PATCHUR COMMUNITY HEALTH CENTRE	Lab No	10642
Constituent Group	Chemistry III	Date of Result Entered :	17/02/2022
PT item	Lyophilized Serum	Date of Report Published :	05/03/2022

Sl. No	Test Name	Method	Instrument	Sample No	Result	Target	Delta	Delta 2	Delta 3	Delta 4	Delta 5
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1724	141.98	8.75	12.43	123.21 mg/dl	-1.51	0.60	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1646	3.08	10.16	0.31	3.51 mg/dl	1.37	0.02	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1650	111.07	7.00	7.77	119.21 mg/dl	1.05	0.38	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

GOVT PATCHUR COMMUNITY HEALTH CENTRE

CLINICAL LABORATORY, GOVT PATCHUR COMMUNITY HEALTH CENTER TIRUPATTUR DT-635853.

PATCHUR

TAMILNADU635854

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Sanjay
22
Mentor, Umicor
T. Primary Health Cent.
Patchur-635 854