

## LAB MONTHLY SUMMARY



Lab Name GOVT PATCHUR  
COMMUNITY HEALTH  
CENTRE Lab No 10642

Month December Year 2021

Constituent  
Group Chemistry III



Date of Result Entered : 01/12/2021

Date of Report Published : 03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1642	206.41	7.45	15.39	222.56 mg/dl	1.05	0.76
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1607	2.26	10.81	0.24	2.24 mg/dl	-0.08	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1671	109.90	7.07	7.77	97.5 mg/dl	-1.60	0.38

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

*D. Mitesh*

Medical Officer,  
\* Patchur Primary Health Cent.  
Patchur-635 854.