

arkray healthcare Pvt. Ltd.

Plot no. 336, 338, 340, Road no. 3, G.I.D.C.,
Sachin 394 230, (Surat), India. Telephone: 0261-239 7712 to 15
Facsimile: 0261-239 7719, URL: <http://www.arkray.com>

Date: 11 August 2022


PERFORMANCE & CALIBRATION CERTIFICATE

Instrument Name : SPOTCHEM EL 1520
Instrument serial no : 11512009
Installed at : Alpha Care Specialty Lab (Gudalur)
Inspection and calibration date: 11/08/2022
Next Due Inspection and calibration date: 10/02/2023

Self-check O.K

- Probe Test: Good
- E Plate Movement OK.
- E Plate Results O.K

On basis of above-mentioned checks, this is to certify that **SPOTCHEM EL 1520** with Serial no **11512009** installed at **Alpha Care Specialty Lab, Gudalur** has been Inspected on **11/08/2022** and performance of the instrument is satisfactory. Calibration in the range. This is valid up to **10/02/2023**, If any major breakdown occurs, instrument must be re-evaluated for performance and calibration.



Arkray HealthCare Pvt. Ltd.
Technical Support Cell

SE-1520 V2.03 2022-02-04 05:25
No.0001 ID:QC
Sample:B. Serum Lot:A2C23

Na	127	mmol/l
K	3.7	mmol/l
Cl	98	mmol/l

Use before 31/07/20

Disclaimer

All laboratory test result must be interest within the context of overall health of the patient and should be used along with other tests and clinical findings. Laboratory test results may vary depending upon age, sex, time of sample collection, diet, medication and physiological variations.