

# MK BEST CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,  
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340

## CERTIFICATE OF CALIBRATION

FF/7.8/01				Page No		1 of 1					
ULR No	CC334022000014530F	Date of Calibration	10.08.2022	Date of Receipt	10.08.2022						
Certificate No	MKBL/22/08/0964-004	Recom. Due Date	09.08.2023	Date of Issue	13.08.2022						
<b>CUSTOMER INFORMATION</b>				<b>DETAILS OF UNIT UNDER CALIBRATION</b>							
M/S ., ALPHIA CARE SPECIALITY LAB , D.J COMPLEX , DEVARSHOLA ROAD , GUDALUR , THE NILGIRIS - 643 212 .				Description		MICROPIPETTE - 2					
				Make / Model		THERMO FISHER / FINN PIPETTE					
				Range/Resolution		5 µl to 50 µl / 1µl					
				Serial No		MW 21061					
				Identification No		ACL/CB / GEN/03					
				Calibrated at		LAB					
<b>STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)</b>											
S.No	Description	Id.No/Sl. No	Certificate No	Validity							
01	Electronic Semi Micro Balance	MK/CAL-96/477904	TVCSPL 22/07/1229	22.07.2023							
<b>ENVIRONMENTAL &amp; DUC CONDITIONS</b>				<b>REFERENCE STANDARD &amp; ACCEPTANCE LIMIT</b>							
Temperature	23 ± 1.5°C		Reference Std	ISO 8655 - 6 : 2002							
Humidity	40 - 60 % RH		Procedure No	MKBCS - MBV - 03							
Condition of DUC Receipt	Good										
<b>CALIBRATION RESULTS</b>											
<b>I. VOLUME CALIBRATION</b>											
S.No	DUC Reading (Mean) µl	STD Reading (Mean) µl	Deviation µl	Expanded Uncertainty (±) µl							
1	10	10.02	-0.02	1.18							
2	20	20.05	-0.05								
3	30	30.08	-0.08								
3	40	40.11	-0.11								
3	50	50.13	-0.13								
<b>Remarks :</b>											
1. The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.											
2. The Calibration Certificate Shall not be Reproduced Expect In Full, Without Written Approval Of The Laboratory.											
3. The Recalibration Interval Should be Determined on the User Requirement.											
4. The Results Stated In This Certificate Relate Only to the Item Calibrated.											
5. The User Should Determine The Suitability Of The Instrument For Is Intended Use.											
6. Resulted Volume Convert at 27°C of Water Temperature.											
7. Expanded Uncertainty is also Included Correction Factors.											
x-x-x-x- End Of Certificate -x-x-x-x											
Calibrated by  S. Murugesan (Calibration Engineer)				 				Authorised By  L. Magesh (MD/QM)			