

GOVT.PRIMARY HEALTH CENTRE-MAPPILAIURANI-JULY-2022-EQAS REPORT

LAB MONTHLY SUMMARY

Lab Name GOVT PRIMARY HEALTH CENTRE Mappilaiurani Lab No 3837
 Month July Year 2022
 Constituent Group Chemistry III

Result Entered Date : 18/07/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1864	295.51	7.14	21.09	289.4 mg/dL	-0.29	0.98
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1839	3.40	9.21	0.31	3.7 mg/dL	0.56	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1803	117.79	6.19	7.29	116 mg/dL	-0.38	0.34

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****