



YEA

☎ : 033 2677 1412, 2667 1909, 82329 85881
Tele Fax : 033 2677 1411

YOUNG ENGG. & CALIBRATION SERVICES PVT. LTD.

Admn. Office & Lab : 44, Shahid Khudiram Bose Sarani, (formerly Komardanga Road), Ichapur, Howrah - 711 104
E-mail : calibration@yeagroup.org ★ Visit us at : www.yeagroup.org
CIN : U74140WB2005PTC106389 ★ Toll Free Help Desk : 1800 345 9001

Calibration Certificate

Certificate No.: YEA(C)/ 220205SA/K652/002
ULR No. : CC222122000006471F

Dated : 08.02.2022

Calibrated for : Kulikawn Hospital, ICTC
Kulikawn,
Aizawl, Pin - 796005.

Calibrated on.	Next Calibration Due On.	Page No.	No. of Pages
05.02.2022	05.02.2023	1	2

Description of Instrument

Name : Timer of VDRL Shaker
Identification : KH/ICTC VDRL-01
Range : 0 to 15 min
Least Count : 1 min
Make : Remi
Model : RS-12R

Service Request No. / Order Reference : 220205SA/K652/001-002

Receipt of Instrument : 05.02.2022

Location where calibration performed : At site

Applicable Specification of Instrument to be Calibrated : Not Specified

Condition of Instrument to be Calibrated : Satisfactory

Basis of Calibration : Comparison method with reference to YEA W1EIM1-06

Environmental Condition during Calibration : Temp. : 25±2°C RH : 50±15%

Equipment used for Calibration : Standard Digital Stop Watch


Traceability : To National Standard vide Certificate No: YEA(C)/211211/Y001/001 valid till 16.12.2022

Measurement Uncertainty : The figures are stated in the result at coverage factor $k = 2$ corresponding to confidence probability of 95% for normal distribution.

Laboratory Identification : Sticker is affixed for authentication of calibration



The above Instrument/Equipment/Machine has been calibrated up to its range.
The results are given in the page 2 to ... 2 ... of the certificate.





YEA

☎ : 033 2677 1412, 2667 1909, 82329 85881

Tele Fax : 033 2677 1411

YOUNG ENGG. & CALIBRATION SERVICES PVT. LTD.

Admn. Office & Lab : 44, Shahid Khudiram Bose Sarani, (formerly Kamardanga Road), Ichapur, Howrah - 711 104

E-mail : calibration@yeagroup.org ★ Visit us at : www.yeagroup.org

CIN : U74140WB2005PTC106389 ★ Toll Free Help Desk : 1800 345 9001

Certificate No.: YEA(C)/ 220205SA/K652/002

ULR No. : CC222122000006471F

Calibrated By : S. Ghosh

Dated : 08.02.2022

Page 2 of 2 Pages

Calibration Results

Electro-Technical Calibration

CALIBRATION RESULTS OF TIMER OF VDRL SHAKER :

Sl. No.	Parameter	DUC Value	Standard Value	Measurement Uncertainty
1	Time	1 min	1 min 0.02 sec	34.64 (± sec)
2		5 min	5 min 0.07 sec	
3		10 min	10 min 0.12 sec	
4		15 min	15 min 0.17 sec	

DUC : Device Under Calibration

REMARKS : The above Timer of VDRL Shaker has been calibrated and the readings are tabulated above.

NOTES : This certificate refers to the results tabulated at the time of calibration only for the above mentioned item submitted for calibration and this should not be reproduced, other than in full except with prior written approval from the head of calibration laboratory.

Authorised By

A. Das

Manager Calibration

*** End of the Calibration Certificate ***

