

MK BEST CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.



Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340

CERTIFICATE OF CALIBRATION

FF/7.8/01		Page No			1 of 1	
ULR No	CC334022000014128F	Date of Calibration	05.08.2022	Date of Receipt	05.08.2022	
Certificate No	MKBL/22/08/0945-004	Recom. Due Date	04.08.2023	Date of Issue	12.08.2022	
CUSTOMER INFORMATION		DETAILS OF UNIT UNDER CALIBRATION				
M/S., SANTHI VALLUVAR LABORATORY, AMBATTUR, CHENNAI - 53		Description	MICROPIPETTE - 2			
		Make / Model	MERILETTE / ND454972			
		Range/Resolution	100 µl to 1000 µl / 5 µl			
		Serial No	VV 100 - 1000			
		Identification No	SVL / CB / GEN / 03			
		Calibrated at	LAB			
STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)						
S.No	Description	Id.No/Sl. No	Certificate No	Validity		
01	Electronic Semi Micro Balance	MK/CAL-96/477904	TVCSPL 22/07/1229	22.07.2023		
ENVIRONMENTAL & DUC CONDITIONS		REFERENCE STANDARD & ACCEPTANCE LIMIT				
Temperature	23 ± 1.5°C	Reference Std	ISO 8655 - 6 : 2002			
Humidity	40 - 60 % RH	Procedure No	MKBCS - MBV - 03			
Condition of DUC Receipt	Good					
CALIBRATION RESULTS						
1.VOLUME CALIBRATION						
S.No	DUC Reading (Mean) µl	STD Reading (Mean) µl	Deviation µl	Expanded Uncertainty (±) µl		
1	100	100.12	-0.12	7.29		
2	300	300.26	-0.26			
3	500	500.37	-0.37			
3	700	700.49	-0.49			
3	1000	1000.63	-0.63			
Remarks :						
1. The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.						
2. The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.						
3. The Recalibration Interval Should be Determined on the User Requirement.						
4. The Results Stated In This Certificate Relate Only to the Item Calibrated.						
5. The User Should Determine The Suitability Of The Instrument For Its Intended Use.						
6. Resulted Volume Convert at 27°C of Water Temperature.						
7. Expanded Uncertainty is also Included Correction Factors.						
x-x-x-x- End Of Certificate -x-x-x-x						
Calibrated by  S. Murugesan (Calibration Engineer)				Authorised By  L. Magesh (MD/QM)		