



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2022



PC-1024

Lab Name GOVT PRIMARY HEALTH CENTRE Lab No 10578
Constituent Group Chemistry III Date of Result Entered : 21/04/2022
PT item Lyophilized Serum Date of Report Published : 05/05/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1781	263.81	7.17	18.93	261.64 mg/dL	-0.11	0.90
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1758	4.67	9.43	0.44	4.35 mg/dL	-0.73	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1741	115.66	6.19	7.16	119.32 mg/dL	0.51	0.34

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
GOVT PRIMARY HEALTH CENTRE
CLINICAL LABORATORY , RAMANATHAPURAM District.,
VENKITANKURICHI
TAMILNADU623707

Coordinator Contact Details:
Email:clinq@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****