



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2022



PC-1024

Lab Name **OSCAR PATH LABS**

Lab No **16081**

Constituent Group **Chemistry II**

Date of Result Entered : **17/05/2022**

PT item **Lyophilized Serum**

Date of Report Published : **08/06/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	167	151.95	9.03	13.72	147 mg/dL	-0.36	2.12
2	UREA II	Urease UV / GLDH II	Mindray	136	120.72	10.57	12.76	146 mg/dL	1.98	2.19
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	104	7.99	12.58	1.00	7.8 mg/dL	-0.19	0.20
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	143	1.74	18.02	0.31	1.2 mg/dL	-1.73	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	159	5.88	8.09	0.48	6.5 g/dL	1.30	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	163	3.56	8.04	0.29	3.5 g/dL	-0.21	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	160	5.64	14.69	0.83	4.7 mg/dL	-1.14	0.13
8	CHOLESTROL II	CHOD-PAP II	Mindray	171	123.02	7.71	9.49	132 mg/dL	0.95	1.45
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	159	105.34	12.07	12.72	105 mg/dL	-0.03	2.02

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

OSCAR PATH LABS

A-1/1, IIIRD FLOOR, ABOVE MG SHOWROOM OUTER RING ROAD,

NEW DELHI

DELHI110085

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****