

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Ramnad Clinical Laboratory, #16C, AKS Complex, R.R. Sethupathy Nagar, Ramanathapuram – 623 501	Certificate No:	AMS/LAB/2889822
		Issuing Date:	25/09/2021
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Incubator	Id. No:	RCL/LAB/INB/001
Make:	LABTEK	Department:	Blood Bank
Model:	Manual	Calibrated On:	24/09/2021
Sl. No:	---	Calibration Due Suggested:	23/09/2022
Type:	Rotary Manual	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 120°C / LC 1°C		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC/01/21/ATL/0107-01	22/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

SL.No.	DUC Reading (°c)	Standard Reading (°c)	Correction (°c)
1	36.0	36.31	+0.31
2	37.0	37.42	+0.42

Allowable Tolerance: ±1°C

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. S. S.



Approved by:

(Technical Manager)

B. S. S.