

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Ramnad Clinical Laboratory, #16C, AKS Complex, R.R. Sethupathy Nagar, Ramanathapuram – 623 501	Certificate No:	AMS/LAB/2889823
		Issuing Date:	25/09/2021
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micro Pipette	Id. No:	RCL/LAB /FMP/001
Make:	ERIC Bio	Department:	Laboratory
Model:	Fixed	Calibrated On:	24/09/2021
Sl. No:	---	Calibration Due Suggested:	23/09/2022
Type:	Fixed	DUC Condition on Receipt:	OK
Range & Least Count:	500µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/21/ATL/0339-01	22/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	500	490	490	-10.0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.



Calibrated By:  
(Calibration Engineer)

B. S. S.



Approved by:  
(Technical Manager)

B. S. S.