



Thrissur Office:

Thattil Nadakalan Complex, Kuriachira P.O.,

Thrissur - 680 006, INDIA.

Tel.: (0487) 2250793, 6544048. Mobile No.: 9645084048.

E-mail: tulipkerala@gmail.com

10.07.2022

CALIBRATION CERTIFICATE

This is to certify that, **EVOLUTION 3000** Semi-Automatic Analyzer bearing Serial No **42229**

Installed at, **M/S MEDLINE DIAGNOSTICS P LTD, NO 23-A, THENNUR HIGH ROAD, TIRUCHIRAPPALLI, TAMILNADU-620017** is performing satisfactorily.

EVOLUTION 3000 is calibrated last on **10.07.2022** by our Service Engineer,

Mr. S. JAWAHAR

Following checks have been performed:

1. Peristaltic Pump Cleaned
2. Cleaned flowcell & Tubings
3. Adjusted Lamp Gain.
4. Checked all the Voltage.

Next Calibration Due on **09.07.2023**

For **TULIP DIAGNOSTICS (P) LTD**

AUTHORISED SIGNATORY

