



**STATE REFERENCE LABORATORY**  
**DEPARTMENT OF MICROBIOLOGY**  
B. R. D. MEDICAL COLLEGE GORAKHPUR-273013

UPS & CS

SRL/5994 /Micro/2021

Date:- 17/12/2021

**External Quality Assessment Scheme for HIV Serology**

**Proficiency Testing Report (SRL-ICTC)**

Name of Participating Laboratory :-			PPTCT – District Women Hospital, Gorakhpur				
EQAS Code	Sample Code	True Status	ICTC Result			Matching YES/NO	% Matching
			Rapid Test #1	Rapid Test #2	Rapid Test #3		
			Combaids	Standard	Tredro		
			40000074369	0057023	MI012046		
			12/2022	12/2022	12/2021		
E731	N1	Reactive	Reactive	Reactive	Reactive	Yes	100%
E733	N2	Non-Reactive	Non-Reactive	---	---		
E734	N3	Non-Reactive	Non-Reactive	---	---		
E732	N4	Reactive	Reactive	Reactive	Reactive		

**Remark: Performance is Satisfactory**

Compiled By:  
Mr. Jyoti Kumar

Technical officer

Approved By:  
Dr. Amresh Kr. Singh

Head  
Department of Microbiology  
B.R.D. Medical College  
Gorakhpur



**STATE REFERENCE LABORATORY**  
**DEPARTMENT OF MICROBIOLOGY**  
B. R. D. MEDICAL COLLEGE GORAKHPUR-273013



SRL/135/Micro/2022

Date:- 31/03/2022

**External Quality Assessment Scheme for HIV Serology**

**Proficiency Testing Report (SRL-ICTC)**

Name of Participating Laboratory :-			PPTCT – District Women Hospital, Gorakhpur				
EQAS Code	Sample Code	True Status	ICTC Result			Matching YES/NO	% Matching
			Rapid Test #1	Rapid Test #2	Rapid Test #3		
			Combaids	WBPFT	Tredro		
			4000024369	MI092149	MI082016		
			12/2022	08/2023	07/2022		
T771	A	Reactive	Reactive	Reactive	Reactive	Yes	100%
T773	B	Non-Reactive	Non-Reactive	---	---		
T774	C	Non-Reactive	Non-Reactive	---	---		
T772	D	Reactive	Reactive	Reactive	Reactive		

**Remark: Performance is Satisfactory**

Compiled By:  
Mr. Jyoti Kumar

**JYOTI KUMAR**  
Technical officer  
SRL, Dept. of Microbiology  
B.R.D. Medical College, Gorakhpur

Approved By:  
Dr. Amresh Kr. Singh

**Head**  
Department of Microbiology  
B.R.D. Medical College  
Gorakhpur