



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2022



Lab Name Bethel X - Ray , ECG & Clinical Laboratory
Constituent Group Chemistry II
PT item Lyophilized Serum

Lab No 11638
Date of Result Entered : 20/07/2022
Date of Report Published : 05/08/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	841	313.52	11.25	35.28	284.6 mg/dL	-0.82	2.43
2	UREA II	Urease UV / GLDH II	Transasia / Erba	698	59.17	10.93	6.46	55.21 mg/dL	-0.61	0.49
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	552	3.38	10.92	0.37	3.51 mg/dL	0.35	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	756	1.82	18.13	0.33	2.16 mg/dL	1.03	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	729	5.98	9.10	0.54	6.04 g/dL	0.11	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	718	3.39	8.38	0.28	4.79 g/dL	5	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	772	7.91	12.86	1.02	8.16 mg/dL	0.25	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	826	116.69	10.53	12.28	137.4 mg/dL	1.69	0.85
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	750	253.80	11.48	29.13	277.4 mg/dL	0.81	2.13

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Bethel X - Ray , ECG & Clinical Laboratory
96 - B , Kanthan Colony Entrance , Palayamkottai Road West , 3rd Mile ,
Tuticorin
TAMILNADU628008

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****