

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610

**AMS**  
Healthcare

GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S.BETHEL CLINICAL LABORATORY, 96-B, Kanthan Colony Entrance, Palayam Kottai road,3 <sup>rd</sup> Mile , Tuticorin-8.	Certificate No:	AMS/HOSP/289/1518
		Issuing Date:	10/06/2022
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Refrigerator - Reagent	Id. No:	BETL/LAB/REFR/001
Make:	SAMSUNG	Department:	Laboratory
Model:	RR19TLY1BU2	Calibrated On:	08/06/2022
Sl. No:	HL-2020	Calibration Due Suggested:	07/06/2023
Type:	Reagent	DUC Condition on Receipt:	OK
Range & Least Count:	-10 to 8°C/ LC 1°C		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC/01/22/ATL/0171-01	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### Test Results:

SL.No.	Place	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	Freezer	FRZ 1	-1.18	0
2		FRZ 2	-0.96	0
3	Chiller	CHI 1	1.13	0
4		CHI 2	1.08	0
5	Chamber I	CHA 1	3.47	0
6		CHA 2	3.51	0
7	Chamber II	CHA 1	4.33	0
8		CHA 2	4.28	0
9	Chamber III	CHA 1	5.13	0
10		CHA 2	5.24	0

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:  
(Calibration Engineer)

*[Signature]*



www.amsmedical.in

Approved by:  
(Technical Manager)

*[Signature]*

93455 56140, 93455 56148