

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S.JEBA CLINICAL LABORATORY, 48, Bolden Puram Street, Near GH, Tiruchendur road, Tuticorin-628003	Certificate No:	AMS/LAB/289/1485
		Issuing Date:	10/06/2022
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	JCL/LAB/VMP/001
Make:	VISON PLUS	Department:	Laboratory
Model:	---	Calibrated On:	08/06/2022
Sl. No:	---	Calibration Due Suggested:	07/03/2023
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	10 – 100µl		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	10.0	10.00	10.00	+00.00
2	20.0	20.00	20.00	+00.00
3	50.0	55.00	55.00	+05.00
4	80.0	78.00	78.00	-08.00
5	100.0	90.00	90.00	-10.00

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

[Signature]



www.amsmedical.in

Approved by:

(Technical Manager)

[Signature]