

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B, MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S.JEBA CLINICAL LABORATORY, 48, Bolden Puram Street, Near GH, Tiruchendur road, Tuticorin-628003	Certificate No:	AMS/LAB/289/1485
		Issuing Date:	10/06/2022
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	JCL/LAB/VMP/002
Make:	Vision plus	Department:	Laboratory
Model:	---	Calibrated On:	08/06/2022
Sl. No:	YE17CA90139512	Calibration Due Suggested:	07/06/2023
Type:	VARIABLE	DUC Condition on Receipt:	OK
Range & Least Count:	10 – 100µl		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	100.0	110.00	110.00	+10.00
2	200.0	210.00	210.00	+10.00
3	500.0	500.00	500.00	+00.00
4	800.0	800.00	800.00	+00.00
5	1000.0	990.00	990.00	-10.00

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)



Approved by:

(Technical Manager)