

HORIBA

Explore the future

HORIBA India Private Ltd.

246, Okhla Industrial

Estate, Phase - III 110

020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2022-2023/100628

04th Oct 2021

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **ABX YUMIZEN H550** bearing serial number: **006YAXH03049** installed at **PATHKIND DIAGNOSTICS PVT. LTD., FARIDABAD** as calibrated on 04 Oct 2021


Calibrator : ABX MINOCAL

Lot No. : CX463

Expiry Date : 05th Nov 2021

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable

Next calibration cycle is due on 03rd Oct 2022.



Shrish Dixit

(Head- Products & Marketing)

For Horiba India Pvt. Ltd.