



Shruti Clinical Laboratory

Dr. Krishna Desai M.D.(Path & Bact.)

203/ 204, Royal Chambers, N. S. Road No. 1, Opp. Club Millennium, Mumbai - 400 049.
2620 7949 / 2625 5488 / 98694 68892 • shrutilab@gmail.com

SCL

TO WHOMSOEVER IT MAY CONCERN

This is bring into your notice that as per the scheme of MHLEQAS clinical pathology (URINE), sample schedule is twice a year. We have received 1st schedule sample for urine in December 2021, for which reports we have submitted.

After December 2021 we have received urine sample in this month ie August 2022, their results we have submitted yesterday.

Request you to consider December 2021 results and do the needful.



Dr Krishna Desai

Laboratory Director.



Result relates only to the sample as received. Investigations have their limitation. Solitary investigation never confirms the final diagnosis of disease, it only helps in diagnosing the disease in correlation to the clinical findings & other investigations.

Please preserve this report carefully, we will not be able to give duplicate report.

*Home Facilities Available**

SHRUTI CLINICAL LAB (586)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210512 and Sample No 01

Report Date : 30/12/2021

Analyte	Instrument	Result Value	Standard Unit
X Specific Gravity	Uriskan 100 plus	1.030	-
# Blood (urine)	Uriskan 100 plus	1+	-
# Nitrite (urine)	Uriskan 100 plus	Present	-

Legend @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

SHRUTI CLINICAL LAB (586)

Outlier And Analyte Summary Report

Result Details For Cycle No 210512 and Sample No 01


Report Date : 30/12/2021

Instrument : Uriskan 100 plus

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5.5,6,6.5
X Specific Gravity	-	1.030	1.015,1.020,1.025
@ Glucose (urine)	-	4+	+2,+3,+4
@ Protein (Urine)	-	2+	+1,+2,Tr
@ Ketone (urine)	-	Negative	Negative
# Blood (urine)	-	1+	
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
# Nitrite (urine)	-	Present	

Legend @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	2
Evaluated Parameters	7
Outlier Parameters (X)	1
EQAS Score Clinical Path - Urine Routine Chemistry	85.71 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report