



October29,2022

*Calibration Certificate for Fully Automated Analyzer
Model: XL640(Sr. No.:640353)*


This is to certify that the Fully Automated Analyzer, Model: XL640(Sr. No.: 640353) installed at REDCLIFFE LIFETECH PVT Ltd, Plotno:144/C, Ground floor,Chandrasekharpur,Bhubaneswar,Odisha-751016 has been calibrated and the calibration value of all filters gains and all the required values are well within limit.

This calibration is valid till October28,2023

For, Transasia Bio-Medicals Ltd.

Shibananda Biswas

Dy. Regional Service Manager



Result Reprint

Report Type : Controls 27-Aug-2022

Consumable: BIORAD L1

Sl #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
1	26491	BIORAD L1	Na	145.50 mmol/l		0	27-Aug-2022 09:35:12	146.747	3.440	136.427 - 157.067
2	26491	BIORAD L1	K	3.60 mmol/l	F,+1SD	0	27-Aug-2022 09:35:12	3.494	0.060	3.314 - 3.674
3	26491	BIORAD L1	Cl	102.20 mmol/l		0	27-Aug-2022 09:35:12	101.837	1.720	96.677 - 106.997
4	26491	BIORAD L1	SGPTD	26.6 U/L	-1SD	183039	27-Aug-2022 09:37:58	27.770	0.950	24.92 - 30.62
5	26491	BIORAD L1	ALPU	111 U/L		183043	27-Aug-2022 09:38:34	107.750	7.210	86.12 - 129.38
6	26491	BIORAD L1	GGT	64.7 U/L		183045	27-Aug-2022 09:38:52	66.300	6.600	46.5 - 86.1
7	26491	BIORAD L1	PHOS	4.03 mg/dl	+1SD	183047	27-Aug-2022 09:39:10	3.791	0.190	3.221 - 4.361
8	26491	BIORAD L1	BID	0.40 mg/dl		183049	27-Aug-2022 09:39:28	0.396	0.020	0.336 - 0.456
9	26491	BIORAD L1	BIT	0.94 mg/dl		183051	27-Aug-2022 09:39:46	0.958	0.080	0.718 - 1.198
10	26491	BIORAD L1	CRENZ	1.72 mg/dl		183055	27-Aug-2022 09:40:22	1.707	0.040	1.587 - 1.827
11	26491	BIORAD L1	GLU	86.1 mg/dl	+1SD	183057	27-Aug-2022 09:40:40	82.853	2.520	75.293 - 90.413
12	26491	BIORAD L1	CHOL	265 mg/dl	+1SD	183061	27-Aug-2022 09:41:16	252.400	7.210	230.77 - 274.03
13	26491	BIORAD L1	UA	5.2 mg/dl		183065	27-Aug-2022 09:41:52	5.194	0.210	4.564 - 5.824
14	26491	BIORAD L1	CA	9.5 mg/dl		183073	27-Aug-2022 09:43:04	9.217	0.370	8.107 - 10.327
15	26491	BIORAD L1	ALB	4.08 g/dl	+1SD	183075	27-Aug-2022 09:43:22	3.965	0.100	3.665 - 4.265
16	26491	BIORAD L1	PRO	6.82 g/dl	+1SD	183077	27-Aug-2022 09:43:41	6.469	0.310	5.539 - 7.399
17	26491	BIORAD L1	SGOTD	35.6 U/L		183087	27-Aug-2022 10:21:47	36.210	1.040	33.09 - 39.33

Print Date 27-Aug-2022



Result Reprint

Report Type : Controls 27-Aug-2022

Consumable: BIORAD L1

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
18	26491	BIORAD L1	UREA	34.4 mg/dl	+1SD	183094	27-Aug-2022 10:22:50	32.681	1.380	28.541 - 36.821
19	26491	BIORAD L1	TRIG	168.1 mg/dl	-1SD	183101	27-Aug-2022 10:23:53	179.873	7.120	158.513 - 201.233
20	26491	BIORAD L1	CHOL	261 mg/dl	+1SD	183102	27-Aug-2022 10:24:03	252.400	7.210	230.77 - 274.03
21	26491	BIORAD L1	HDLC	59.6 mg/dl	+1SD	183103	27-Aug-2022 10:24:11	57.243	2.020	51.183 - 63.303



Result Reprint

Report Type : Controls 27-Aug-2022 Consumable: BIORAD L2

Sl#	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
1	89702	BIORAD L2	Na	125.10 mmol/l		0	27-Aug-2022 09:35:59	124.000	6.250	105.25 - 142.75
2	89702	BIORAD L2	K	5.66 mmol/l	F	0	27-Aug-2022 09:35:59	5.680	0.280	4.84 - 6.52
3	89702	BIORAD L2	Cl	86.80 mmol/l		0	27-Aug-2022 09:35:59	85.900	4.350	72.85 - 98.95
4	89702	BIORAD L2	SGPTD	93.2 U/L		183040	27-Aug-2022 09:38:06	94.700	9.550	66.05 - 123.35
5	89702	BIORAD L2	SGOTD	186.1 U/L		183042	27-Aug-2022 09:38:25	187.000	18.000	133 - 241
6	89702	BIORAD L2	ALPU	500 U/L		183044	27-Aug-2022 09:38:43	517.000	51.000	364 - 670
7	89702	BIORAD L2	GGT	171.9 U/L		183046	27-Aug-2022 09:39:01	176.000	18.000	122 - 230
8	89702	BIORAD L2	PHOS	7.45 mg/dl		183048	27-Aug-2022 09:39:19	7.500	0.600	5.7 - 9.3
9	89702	BIORAD L2	BID	1.09 mg/dl	-1SD	183050	27-Aug-2022 09:39:37	1.280	0.160	0.8 - 1.76
10	89702	BIORAD L2	BIT	4.03 mg/dl		183052	27-Aug-2022 09:39:55	4.260	0.540	2.64 - 5.88
11	89702	BIORAD L2	UREA	112.0 mg/dl	+1SD	183054	27-Aug-2022 09:40:13	100.000	7.500	77.5 - 122.5
12	89702	BIORAD L2	CRENZ	5.10 mg/dl		183056	27-Aug-2022 09:40:31	5.200	0.450	3.85 - 6.55
13	89702	BIORAD L2	GLU	295.9 mg/dl		183058	27-Aug-2022 09:40:49	291.000	22.500	223.5 - 358.5
14	89702	BIORAD L2	TRIG	83.5 mg/dl		183060	27-Aug-2022 09:41:07	85.700	6.450	66.35 - 105.05
15	89702	BIORAD L2	CHOL	103 mg/dl		183062	27-Aug-2022 09:41:25	101.000	7.750	77.75 - 124.25
16	89702	BIORAD L2	HDLc	21.2 mg/dl		183064	27-Aug-2022 09:41:43	23.000	2.250	16.25 - 29.75
17	89702	BIORAD L2	UA	10.1 mg/dl		183066	27-Aug-2022 09:42:01	10.400	0.750	8.15 - 12.65

Print Date 27-Aug-2022

Result Reprint

Report Type : Controls 27-Aug-2022 Consumable: BIORAD L2

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
18	89702	BIORAD L2	CA	12.6 mg/dl		183074	27-Aug-2022 09:43:13	13.000	1.050	9.85 - 16.15
19	89702	BIORAD L2	ALB	2.71 g/dl		183076	27-Aug-2022 09:43:31	2.700	0.210	2.07 - 3.33
20	89702	BIORAD L2	PRO	4.71 g/dl		183078	27-Aug-2022 09:43:49	4.600	0.300	3.7 - 5.5

Print Date 27-Aug-2022

