



October29,2022

*Calibration Certificate for Urine Analyzer
Model: Laura Smart(Sr. No.:6813-1-202102)*

This is to certify that the Urine Analyzer, Model: Laura Smart (Sr. No.: 6813-1-202102) installed at REDCLIFFE LIFETECH PVT Ltd, Plotno:144/C, Ground floor, Chandrasekharpur, Bhubaneswar, Odisha-751016 has been calibrated and the calibration value of all filters gains and all the required values are well within limit.

This calibration is valid till October28,2023

For, Transasia Bio-Medicals Ltd.

Shibananda Biswas

Dy. Regional Service Manager



LAURA
S M A R T

.....
Laura Smart
Version 5.0 Ma
.....
Selftest OK.
27.08.2022 10:53



LAURA
S M A R T

DekaPHAN LAURA

Seq.No:0001
ID:URINORM-N
27.08.2022 11:13

.....
BLD NEG
LEU NEG
BIL NEG
UBG NORM
KET NEG
GLU NEG
PRO NEG
PH 6
NIT NEG
SG 1.030

LAURA
S M A R T

DekaPHAN LAURA

Seq.No:0003
ID:URINORM-P
27.08.2022 11:18



.....
*BLD 250 Ery/ μ l
*LEU 500 Leu/ μ l
BIL NEG
UBG NORM
*KET 52 mg/dl
*GLU 250 mg/dl
*PRO 100 mg/dl
PH 9
NIT NEG
SG 1.000