



October 29, 2021

*Calibration Certificate for Urine Analyzer  
Model: Laura Smart (Sr. No.: 6813-1-202102)*

*This is to certify that the Urine Analyzer, Model: Laura Smart (Sr. No.: 6813-1-202102) installed at REDCLIFFE LIFETECH PVT Ltd, Plotno: 144/C, Ground floor, Chandrasekharapur, Bhubaneswar, Odisha-751016 has been calibrated and the calibration value of all filters gains and all the required values are well within limit.*

*This calibration is valid till October 28, 2022*

*For, Transasia Bio-Medicals Ltd.*

*Shibananda Biswas*

*Dy. Regional Service Manager*



**LAURA**  
S M A R T

.....  
Laura Smart  
Version 5.0 Ma  
.....  
Selftest OK.  
27.08.2022 10:53



**LAURA**  
S M A R T

DekaPHAN LAURA

Seq.No:0001  
ID:URINORM-N  
27.08.2022 11:13  
.....

BLD NEG  
LEU NEG  
BIL NEG  
UBG NORM  
KET NEG  
GLU NEG  
PRO NEG  
PH 6  
NIT NEG  
SG 1.030

**LAURA**  
S M A R T

DekaPHAN LAURA

Seq.No:0003  
ID:URINORM-P  
27.08.2022 11:18  
.....



\*BLD 250 Ery/ $\mu$ l  
\*LEU 500 Leu/ $\mu$ l  
BIL NEG  
UBG NORM  
\*KET 52 mg/dl  
\*GLU 250 mg/dl  
\*PRO 100 mg/dl  
PH 9  
NIT NEG  
SG 1.000