


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 220819A/N-113	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 20 / 08 / 2022	<b>Ref. Std.</b>	: CG-11
<b>CALIBRATION DUE.</b>	: 19 / 08 / 2023	<b>C.P. No.</b>	: ACS/CP/07
<b>SRF No. &amp; DATE</b>	: 220819A/N Dt. 19.08.2022	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342722000000363F Dt. 25.08.2022	<b>Cal. At</b>	: Site / Lab

<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. Plot No. 144/C, Ground Floor, Chandrasekharpur, Bhubaneswar, Distt. Khordha-751016, Odisha.
-----------------------	--

<b>ITEM CALIBRATED</b>	: Digital Temp. Indicator of Refrigerator		
<b>Make</b>	: Apteckdeals	<b>Range</b>	: (-50 to 110) °C
<b>L. Count</b>	:	<b>ID No.</b>	: RCL/HEM/TEMP-01
<b>Location</b>	: Haematology Fridge	<b>S. No.</b>	: N.A

<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C ± 3°C } R.H. {40% to 60%}
-------------------------------	--

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2021/0784	11/11/2022

RESULTS			ET Discipline
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)	
2.04	2	0.04	
4.08	4	0.08	
6.12	6	0.12	
8.16	8	0.16	
(average of readings)			

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

