



October 29, 2021

*Calibration Certificate for Urine Analyzer
Model: Laura Smart (Sr. No.: 6813-1-202102)*

This is to certify that the Urine Analyzer, Model: Laura Smart (Sr. No.: 6813-1-202102) installed at REDCLIFFE LIFETECH PVT Ltd, Plotno: 144/C, Ground floor, Chandrasekharapur, Bhubaneswar, Odisha-751016 has been calibrated and the calibration value of all filters gains and all the required values are well within limit.

This calibration is valid till October 28, 2022

For, Transasia Bio-Medicals Ltd.

Shibananda Biswas

Dy. Regional Service Manager



LAURA[®]
S M A R T

DekaPHAN LAURA

ID: URINORM-N
30.08.2022 10:36

.....
BLD NEG
LEU NEG
BIL NEG
UBG NORM
KET NEG
GLU NEG
PRO NEG
pH 6
NIT NEG
SG 1.025

QC TEST OK



LAURA[®]
S M A R T

DekaPHAN LAURA

ID: URINORM-P
30.08.2022 10:42

.....
*BLD 250 Ery/ μ l
*LEU 500 Leu/ μ l
*BIL 1 mg/dl
*UBG 1 mg/dl
*KET 52 mg/dl
*GLU 500 mg/dl
*PRO 100 mg/dl
pH 9
*NIT POS
SG 1.000

QC TEST OK

