

Calibration Certificate

ULR No.:	CC309622000001726F		
Certificate No.:	CCTPL/MP/0197/05	Field:	Mechanical
SRF No.:	190/22	Date of Receipt:	26/04/2022
SRF Date:	26/04/2022	Date of Calibration:	26/04/2022
Company Name & Address	M/s. GOLA PATHLABS. A-4 / 252, Sector - 28, Rohini (Opp Baghban Apartment) Delhi- 110042	Recommended Due Date.:	25/04/2023
		(As per agreed by customer)	
		Certificate Issue Date:	26/04/2022

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	-----
Make:	Labmaie	Instrument Id No.:	MP-01
Model No.:	---	Location:	Lab
Range:	5 to 50 µl	DUC Condition:	Ok
Readability:	0.5 µl	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/12080/01	PMM, Noida	06/05/2022

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

CALIBRATION RESULT


Sr. No.	DUC* Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	5	4.96	0.04	4.97
2	25	24.89	0.11	24.92
3	50	49.82	0.18	49.88

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By : 
Nishant Kumar
Calibration Engineer



Approved By : 
Sandeep Sharma
Director

- Note. 1.This Certificate only to the particular instrument submitted for calibration.
2.This calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL

---End of Certificate---

Page 1 of 1