

### Calibration Certificate

ULR No.:	CC309622000001722F		
Certificate No.:	CCTPL/CF/0197/02	Field:	Mechanical
SRF No.:	190/22	Date of Receipt:	26/04/2022
SRF Date:	26/04/2022	Date of Calibration:	26/04/2022
Company Name & Address	M/s. GOLA PATHLABS A-4 / 252, Sector - 28, Rohini (Opp Baghban Apartment) Delhi- 110042	Recommended Due Date.:	25/04/2023
		(As per agreed by customer)	
		Certificate Issue Date:	26/04/2022

**DUC\* Details**

Instrument Name:	Centrifuge	Instrument Sr. No.:	ZHCN-11165
Make:	----	Instrument Id No.:	CF-18
Model No.:	----	Location:	Lab
Range:	0 to 3500 rpm	DUC Condition:	Ok
Readability:	100 rpm	Calibration Performed at:	At Site

**Standard Equipments Used(Traceable to National Standard)**

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Digital Tachometer	PMM/13264/1	PMMPL	27/08/2022

Environment Condition	Temperature	Humidity	Reference	Calibration Procedure No.
		25±5°C	55±10%RH	Comparison Method

**Calibration Result**

Sr. No.	Standrad Reading (rpm)	Duc Reading (rpm)	Error (rpm)	Uncertainty %rdg
1	1000.4	1000	-0.4	1.6
2	2000.8	2000	-0.8	1.6
3	3001.2	3000	-1.2	1.6
4	3501.5	3500	-1.5	1.6

Note : DUC\* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2,

Calibrated By : *Nishant Kumar*  
Calibration Engineer



Approved By : *Sandeep Sharma*  
Director

Note. 1 This Certificate only to the particular instrument submitted for calibration.  
2 This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.  
3. This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

—End of Certificate—

Page 1 of 1