

Calibration Certificate

ULR No.:	CC309622000001721F		
Certificate No.:	CCTPL/IN/0197/01	Field:	Thermal
SRF No.:	190/22	Date of Receipt:	26/04/2022
SRF Date:	26/04/2022	Date of Calibration:	26/04/2022
Company Name & Address	M/s. GOLA PATHILABS A-4 / 252, Sector - 28, Rohini (Opp Baghban Appartment) Delhi- 110042	Recommended Due Date.:	25/04/2023
		(As per agreed by customer)	
		Certificate Issue Date:	26/04/2022

DUC* Details

Instrument Name:	Incubator	Instrument Sr. No.:	----
Make	----	Instrument ID	IB-01
Model No.	----	Location:	-----
Range:	0 to 70 °C	DUC Condition:	OK
Readability:	0.1 °C	Calibration Performed at:	At Site

Standard Equipments Used (Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	SPRT With Super DAQ	TL/021/1001.1.1	Tempsens	01/10/2022

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
		25±5°C	55±10%RH	DKD-R-5-7

Calibration Result

Sr. No.	Nominal Temperature (°C)	Standard Reading (°C)	DUC* Reading (°C)	Deviation (°C)	Expanded Uncertainty ±(°C)
1	20	19.985	20	0.015	1.0
2	30	29.956	30	0.044	1.0
3	50	49.935	50	0.065	1.0
4	60	59.865	60	0.135	1.0
5	70	69.823	70	0.177	1.0

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2.

Calibrated By :
Nishant Kumar
Calibration Engineer



Approved By :
Sandeep Sharma
Director

Note. 1.This Certificate only to the particular instrument submitted for calibration.
2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---

Page 1 of 1