



### Calibration Certificate

ULR No.:	CC309622000001829F		
Certificate No.:	CCTPL/DTI/0223/01	Field:	Mechanical
SRF No.:	216/22	Date of Receipt:	02/05/2022
SRF Date:	02/05/2022	Date of Calibration:	02/05/2022
Company Name & Address	M/s. GOLA PATHLABS A-4 / 252, Sector - 28, Rohini (Opp Baghban Apartment) Delhi- 110042	Recommended Due Date.:	01/05/2023
		(As per agreed by customer)	
		Certificate Issue Date:	03/05/2022

DUC\* Details

Instrument Name:	DTI With Sensor (Freezer)	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	FD-01
Model No.:	----	Location:	Lab
Range:	-10 to 0 °C	DUC Condition:	OK
Readability:	0.1 °C	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	SPRT With Super DAQ	TL/021/1001.1.1	Tempens	01/10/2022


Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
	25±5°C	55±10%RH	DKD-R-5-7	CCTPL/TH/CP/05S

Calibration Result

Sr. No.	Nominal Temperature (°C)	Standard Reading (°C)	DUC* Reading (°C)	Deviation (°C)	Expanded Uncertainty ±(°C)
1	-2	-2.012	-2.0	0.012	1.0
2	-4	-4.035	-4.0	0.035	1.0
3	-6	-6.085	-6.0	0.085	1.0
4	-8	-8.125	-8.0	0.125	1.0
5	-10	-10.165	-10.0	0.165	1.0

Note : DUC\* : Device Under Calibration  
 Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2.

Calibrated By :   
 Nishant Kumar  
 Calibration Engineer



Approved By :   
 Sandeep Sharma  
 Director

Note 1. This Certificate only to the particular instrument submitted for calibration.  
 2. This Calibration Result reported in this certificate valid for the time of and under the stated condition of measurement.  
 3. This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---