



arkray healthcare Pvt. Ltd.

Plot no. 336, 338, 340, Road no. 3, G.I.D.C.,
Sachin 394 230, (Surat), India. Telephone: 0261-239 7712 to 15
Facsimile: 0261-239 7719, URL <http://www.arkray.co.in>

Certificate Of Calibration

This is to certify that our service engineer Mr. Sachin Singh has attended **U120 Urine analyzer** (Sr.no: **197K10007EE**) on date **28/08/2021** installed at **REDCLIFFE LIFETECH PVT. LTD. Indore-452016**

During his visit, he has **done Calibration & checked performance** of the instrument as listed below

U120 Urine analyzer Serial no. 197K10007EE
Self-check O.K

Main control unit	Pass
Optical sensor	Pass
Test accuracy	Pass
Mechanism	Pass
Excess Light	Pass

On basis of above mention checks, the instrument is functioning satisfactory.
This is valid up to 27/08/2022. If any major breakdown occur instrument to be re-validated.

Best Regards,

Sachin Singh (SE)
Arkray Healthcare Pvt Ltd

Regd. Office: 701/702, Opulence.6th Road, TPS III, Santacruz(East) Mumbai-400 055 INDIA :telephone: +91 - 22-6155 9355 :
facsimile: +91-22-61559356: CIN NO :U51507MH2007PTC175112