



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JANUARY 2022



Certificate No. 17, 1121

Lab Name THE PATHOCARE LABORATORY Lab No 15206
 Constituent Group Chemistry II Date of Result Entered : 12/01/2022
 PT item Lyophilized Serum Date of Report Published : 07/02/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	135	131.73	6.83	8.99	127.6 mg/dl	-0.46	1.55
2	UREA II	Urease UV / GLDH II	Mindray	108	47.21	8.30	3.92	43.1 mg/dl	-1.05	0.75
3	CREATININE II	Enzymatic Colorimetric II	Mindray	33	1.72	6.91	0.12	1.64 mg/dl	-0.67	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	126	2.34	15.93	0.37	2.25 mg/dl	-0.24	0.07
5	Γ-PROTEIN II	Biuret - Colorimetric II	Mindray	124	5.07	8.51	0.43	4.9 g/dl	-0.39	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	137	3.15	9.65	0.30	2.9 g/dl	-0.82	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	125	7.02	10.31	0.72	6.73 mg/dl	-0.40	0.13
8	CHOLESTROL II	CHOD-PAP II	Mindray	132	103.53	8.62	8.92	101.4 mg/dl	-0.24	1.55
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	123	133.64	10.43	13.94	154.1 mg/dl	1.47	2.51

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

THE PATHOCARE LABORATORY
 A-11 WAMINARAYAN AVENUE, MR. AEC CROSS ROAD, 132FT RING ROAD, NARANPURU
 AHMEDABAD
 GUJARAT380013

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

An
Sathishkumar
12/2/2022



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - FEBRUARY 2022



PC-1024

Lab Name THE PATHOCARE LABORATORY Lab No 15206
Constituent Group Chemistry II Date of Result Entered : 12/02/2022
PT item Lyophilized Serum Date of Report Published : 05/03/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	136	273.18	9.97	27.25	250.8 mg/dL	-0.82	4.67
2	UREA II	Urease UV / GLDH II	Mindray	116	29.30	10.78	3.16	33.28 mg/dL	1.26	0.59
3	CREATININE II	Enzymatic Colorimetric II	Mindray	34	1.08	10.71	0.12	0.98 mg/dL	-0.86	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	124	5.16	11.42	0.59	5.07 mg/dL	-0.15	0.11
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	132	5.52	9.97	0.55	5.3 g/dL	-0.40	0.10
6	ALBUMIN II	BCG - Colorimetric II	Mindray	133	3.38	9.28	0.31	3.2 g/dL	-0.57	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	101	4.47	14.61	0.65	3.92 mg/dL	-0.84	0.13
8	CHOLESTROL II	CHOD-PAP II	Mindray	132	115.45	7.89	9.11	113.1 mg/dL	-0.26	1.59
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	138	105.49	12.99	13.70	117 mg/dL	0.84	2.33

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

THE PATHOCARE LABORATORY
A-112, SWAMINARAYAN AVENUE, MR. AEC CROSS ROAD, 132FT RING ROAD, NARANPURU
AHMEDABAD
GUJARAT380013

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Arjun
Sachin
14/03/2022



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2022



PC-1024

Lab Name **THE PATHOCARE LABORATORY** Lab No **15206**
 Constituent Group **Chemistry II** Date of Result Entered : **14/03/2022**
 PT item **Lyophilized Serum** Date of Report Published : **06/04/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	154	107.82	9.16	9.88	105 mg/dL	-0.29	1.59
2	UREA II	Urease UV / GLDH II	Mindray	123	81.80	8.15	6.67	81.94 mg/dL	0.02	1.20
3	CREATININE II	Enzymatic Colorimetric II	Mindray	36	4.20	7.66	0.32	4.18 mg/dL	-0.06	0.11
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	115	1.56	17.13	0.27	1.55 mg/dL	-0.04	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	138	5.13	9.43	0.48	4.6 g/dL	-1.10	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	145	3.13	9.14	0.29	2.9 g/dL	-0.80	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	136	7.72	9.29	0.72	7.41 mg/dL	-0.43	0.12
8	CHOLESTROL II	CHOD-PAP II	Mindray	150	105.86	9.18	9.72	99.8 mg/dL	-0.62	1.59
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	144	202.29	9.46	19.14	203.7 mg/dL	0.07	3.19

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

THE PATHOCARE LABORATORY
A-112, SWAMINARAYAN AVENUE, MR. AEC CROSS ROAD, 132FT RING ROAD, NARANPURU
AHMEDABAD
GUJARAT380013

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Sabzibhushy
14/4/2022



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2022



PC-1024

Lab Name **THE PATHOCARE LABORATORY** Lab No **15206**
Constituent Group **Chemistry II** Date of Result Entered : **14/04/2022**
PT item **Lyophilized Serum** Date of Report Published : **05/05/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	160	219.55	8.60	18.88	206.1 mg/dL	-0.71	2.99
2	UREA II	Urease UV / GLDH II	Mindray	128	37.19	10.60	3.94	37.6 mg/dL	0.10	0.70
3	CREATININE II	Enzymatic Colorimetric II	Mindray	44	5.20	10.85	0.56	5.3 mg/dL	0.18	0.17
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	137	3.01	14.97	0.45	2.95 mg/dL	-0.13	0.08
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	146	5.53	8.58	0.47	4.9 g/dL	-1.33	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	149	3.39	9.02	0.31	3.1 g/dL	-0.95	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	152	6.78	11.82	0.80	6.05 mg/dL	-0.91	0.13
8	CHOLESTROL II	CHOD-PAP II	Mindray	158	112.79	8.29	9.35	105.3 mg/dL	-0.80	1.49
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	153	129.85	10.89	14.14	139.1 mg/dL	0.65	2.29

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

THE PATHOCARE LABORATORY

A-112, SWAMINARAYAN AVENUE, MR. AEC CROSS ROAD, 132FT RING ROAD, NARANPURU

AHMEDABAD

GUJARAT380013

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Am... 12/05/2022
Sab... 12/05/2022



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2022



PC-1024

Lab Name THE PATHOCARE LABORATORY Lab No 15206
 Constituent Group Chemistry II Date of Result Entered : 12/05/2022
 PT Item Lyophilized Serum Date of Report Published : 08/06/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	167	151.95	9.03	13.72	147.5 mg/dL	-0.32	2.12
2	UREA II	Urease UV / GLDH II	Mindray	136	120.72	10.57	12.76	122.18 mg/dL	0.11	2.19
3	CREATININE II	Enzymatic Colorimetric II	Mindray	37	8.81	11.38	1.00	9.29 mg/dL	0.48	0.33
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	143	1.74	18.02	0.31	1.77 mg/dL	0.10	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	159	5.88	8.09	0.48	5.7 g/dL	-0.38	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	163	3.56	8.04	0.29	3.1 g/dL	-1.61	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	160	5.64	14.69	0.83	5.35 mg/dL	-0.35	0.13
8	CHOLESTROL II	CHOD-PAP II	Mindray	171	123.02	7.71	9.49	126.5 mg/dL	0.37	1.45
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	159	105.34	12.07	12.72	114.8 mg/dL	0.74	2.02

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

THE PATHOCARE LABORATORY
 A-112, SWAMINARAYAN AVENUE, MR. AEC CROSS ROAD, 132FT RING ROAD, NARANPURU
 AHMEDABAD

checked
OK
13/06/2022



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2022



PC-1024

Lab Name THE PATHOCARE LABORATORY

Lab No

15206

Constituent Group Chemistry II

Date of Result Entered :

19/07/2022

PT Item Lyophilized Serum

Date of Report Published :

05/08/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD				
1	GLUCOSE II	GOD-POD II	Mindray	156	312.51	8.99	28.09	302.2 mg/dL	-0.37	4.50	
2	UREA II	Urease UV / GLDH II	Mindray	142	59.58	10.77	6.42	62.55 mg/dL	0.46	1.08	
3	CREATININE II	Enzymatic Colorimetric II	Mindray	41	3.49	7.76	0.27	3.66 mg/dL	0.63	0.08	
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	146	1.82	17.98	0.33	1.68 mg/dL	-0.43	0.05	
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	157	5.75	9.39	0.54	5.4 g/dL	-0.65	0.09	
6	ALBUMIN II	BCG - Colorimetric II	Mindray	156	3.39	8.17	0.28	3.4 g/dL	0.04	0.04	
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	156	7.72	10.29	0.80	7.63 mg/dL	-0.11	0.13	
8	CHOLESTROL II	CHOD-PAP II	Mindray	163	119.51	7.82	9.35	120.4 mg/dL	0.10	1.46	
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	147	274.96	7.84	21.56	284.3 mg/dL	0.43	3.56	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

THE PATHOCARE LABORATORY

A-112, SWAMINARAYAN AVENUE, MR. AEC CROSS ROAD, 132FT RING ROAD, NARANPURU
AHMEDABAD

Substanty
Ar
17/8/2022