



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2022



PC-1024

Lab Name	REDCLIFFE LIFETECH PVT LTD	Lab No	15556
Constituent Group	HbA1c	Date of Result Entered :	03/06/2022
PT item	Lyophilized Whole blood	Date of Report Published :	07/07/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	Turbidimetric Inhibition Immunoassay	Any Analyser	279	7.61	11.27	0.86	6.2 %	-1.64	0.10

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

REDCLIFFE LIFETECH PVT LTD  
 FIRST FLOOR, ABOVE SHRI SAI STEEL, NEAR ANUGRAH RESIDENCY, NAHAR ROAD, GONDWARA  
 RAIPUR  
 CHHATTISGARH493221

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*