

MK BEST CALIBRATION SERVICES



एमके बेस्ट केलिब्रेशन सर्विसेस

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CERTIFICATE OF CALIBRATION



FF/7.8/01		Page No		1 of 1	
SRF No	MKBN/22/08/0800	Date of Calibration	30.08.2022	Date of Receipt	30.08.2022
Certificate No	MKBN/22/08/0800-002	Recom. Due Date	29.08.2023	Date of Issue	30.08.2022

CUSTOMER INFORMATION		DETAILS OF UNIT UNDER CALIBRATION	
M/S., SRI VIDHYA DIAGNOSTIC CENTRE, SAIDAPET, CHENNAI - 600 015		Description	MICROPIPETTE - 2
		Make / Model	DRAGON LAB / YEA19AD0065498
		Range/Resolution	100 µl to 1000 µl / 5µl
		Serial No	VV100-1000
		Identification No	SVDC/CB/GEN-03
		Calibrated at	LAB

STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)				
S.No	Description	Id.No/Sl. No	Certificate No	Validity
01	Electronic Semi Micro Balance	MK/CAL-96/477904	TVCSP 22/07/1229	22.07.2023

ENVIRONMENTAL & DUC CONDITIONS		REFERENCE STANDARD & ACCEPTANCE LIMIT	
Temperature	23 ± 1.5°C	Reference Std	ISO 8655 - 6 : 2002
Humidity	40 - 60 % RH	Procedure No	MKBCS - MBV - 03
Condition of DUC Receipt	Good		

CALIBRATION RESULTS				
1.VOLUME CALIBRATION				
S.No	DUC Reading (Mean) µl	STD Reading (Mean) µl	Deviation µl	Expanded Uncertainty (±) µl
1	100	100.62	-0.62	7.29
2	300	300.42	-0.42	
3	500	500.62	-0.62	
4	700	700.22	-0.22	
5	900	900.28	-0.28	
6	1000	1000.34	-0.34	

- Remarks :**
- The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.
 - The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
 - The Recalibration Interval Should be Determined on the User Requirement.
 - The Results Stated In This Certificate Relate Only to the Item Calibrated.
 - The User Should Determine The Suitability Of The Instrument For Its Intended Use.
 - Resulted Volume Convert at 27°C of Water Temperature.
 - Expanded Uncertainty is also Included Correction Factors.

x-x-x-x- End Of Certificate -x-x-x-x

Calibrated by

S. Murugesan
(Calibration Engineer)

Authorised By

L. Magesh
(MD/QM)