


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



CC-3427

Calibration Certificate

CERTIFICATE NO.	: 220826A/N-101	Page	: 1 of 1
CALIBRATED ON.	: 26 / 08 / 2022	Ref. Std.	: ISO:8655:6
CALIBRATION DUE.	: 25 / 02 / 2023	C.P. No.	: ACS/CP/08
SRF No. & DATE	: 220826A/N Dt. 26.08.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342722000000436F Dt. 29.08.2022	Cal. At	: Site / Lab

CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 2 nd Floor Above Shri Sai Steel, Near Anugarh Residency, Nahar Road Gondwara, Raipur (C.G) 493221
-----------------------	---

ITEM CALIBRATED	: Micro Pipette
Make	: Fine Care Corporation
Range	: (100 to 1000) µl
L. Count	:
ID No.	: RCL/BIO/PIP-01
Location	: Biochemistry
S. No.	: PL572221

ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. { 40% to 60% }
-------------------------------	--

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

RESULTS :				
Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		100	100.33	(-) 0.33
1000		500	501.59	(-) 1.59
		1000	1003.22	(-) 3.22

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.12 µl at coverage factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

Auth. Signatory
S.K. Sharma
CEO/TM
C-23, Sec-10, Noida