


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 220826A/N-102	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 26 / 08 / 2022	<b>Ref. Std.</b>	: ISO:8655:6
<b>CALIBRATION DUE.</b>	: 25 / 02 / 2023	<b>C.P. No.</b>	: ACS/CP/08
<b>SRF No. &amp; DATE</b>	: 220826A/N Dt. 26.08.2022	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342722000000437F Dt. 29.08.2022	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. 2 <sup>nd</sup> Floor Above Shri Sai Steel, Near Anugarh Residency, Nahar Road Gondwara, Raipur (C.G) 493221		

<b>ITEM CALIBRATED</b>	: Micro Pipette			
	<b>Make</b>	: Fine Care Corporation	<b>Range</b>	: (2 to 20) µl
	<b>L. Count</b>	:	<b>ID No.</b>	: RCL/BIO/PIP-02
	<b>Location</b>	: Biochemistry	<b>S. No.</b>	: QF599322

**ENVIRONMENT CONDITION:** Temp. { 25°C ± 3°C } R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023 -

RESULTS :				
Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	10.07	(-) 0.07
20		20	20.20	(-) 0.20

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.02 µl at coverage factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

