


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com **RF-3427**



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 220826A/N-108	<b>Page</b> : 1 of 1
<b>CALIBRATED ON.</b> : 27 / 08 / 2022	<b>Ref. Std.</b> : IS:12508:1988
<b>CALIBRATION DUE.</b> : 26 / 08 / 2023	<b>C.P. No.</b> : ACS/CP/10
<b>SRF No. &amp; DATE</b> : 220826A/N Dt. 26.08.2022	<b>Phy. Inspection</b> : O.K.
<b>ULR No. &amp; ISSUE DATE</b> : CC342722000000443F Dt. 29.08.2022	<b>Cal. At</b> : Site / Lab

<b>CALIBRATED FOR</b> : M/s Redcliffe Lifetech Pvt. Ltd. 2 <sup>nd</sup> Floor Above Shri Sai Steel, Near Anugarh Residency, Nahar Road Gondwara, Raipur (C.G) 493221
---

<b>ITEM CALIBRATED</b> : Centrifuge	
<b>Make</b> : Remi	<b>Range</b> : (0 to 5250) RPM
<b>L. Count</b> :	<b>ID No.</b> : RCL/CP/Cent-02
<b>Location</b> : Clinical Pathology	<b>S. No.</b> : ZHFN-18150

<b>ENVIRONMENT CONDITION:</b> Temp. { 25°C ± 3°C } R.H. { 40% to 60% }
--

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ TM5010EK-039	Godrej	M-210914-16-1	14.09.2022

RESULTS : Mechanical Discipline		
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)
1000	1001.68	- 1.68
2000	2002.52	- 2.52
3000	3003.36	- 3.36
4000	4005.74	- 5.74
(average of readings)		

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

