



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276302207 / 9028888728



CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER :-
SHREE DIAGNOSTIC CENTRE
Shree Hospital, Murbad Road, KALYAN

Page No. :- 1 of 1
Certificate No. :- GTS/220205/01-001
Date of Received :- 05.02.2022
Date of Calibration :- 05.02.2022
Next Calibration Due On :- 04.02.2023
Issue Date :- 10.02.2022
Calibration method No. :- TH-WI-02
ULR No :- CC295722000000869F

Ambient Temp. (°C) :- 30.0
Relative Humidity (%RH) :- 51
Location of calibration :- On Site
Condition of Item :- Ok

2. Description of Item

Name :- TIWS(Dry Bath Incubator) Range :- 10 TO 50 °C
Id No :- SDC/DBINC/01 Least Count :- 0.1 °C
Make :- SPAN Location :- Lab
Type :- Digital Sr No IC 01176
Dept Pathology

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2106/002/004	Nishitronics Instrumentation	GTS/PTH/01	31.05.2022

*Thermal Calibration

4.Calibration Results :-

Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
30	30.11	30	-0.11	0.80
37	37.18	37	-0.18	0.80

- Note:**
- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95.45% for normal distribution
 - 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
 - 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
 - 4) Calibration point were selected as per customer specifications.
 - 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
 - 6) UUC :- diameter : 3 mm & immersion depth : 40 mm

Calibrated By
Rohit
Calibration Engineer
Rohit.L



Approved By
Sw
Technical Manager
Swapnil Bhagwat

End of Certificate

RF-55/01