

Transasia Bio-Medicals Ltd , Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai- 400 072
Tel. +91 22 4030 9000 Fax. +91 22 2857 3030

Date: 13/05/2022

Certificate of Calibration

Customer Name: Redcliffe Lifetech Pvt. Ltd. Meerut

Model: Semi Automated Urine Strip Analyzer Erba Laura V2

Serial No.: 0754-1-202109

Calibration done date 13/05/2022

Next Calibration due Date: On or Before 13/05/2023

Lab In charge: Dr. Renu Kamboj

*This is to certify that above mentioned product has been verified of calibration
According to the standard procedures provided by Erba.*

Calibration at site performed by Sharad Srivastava



(Sharad Srivastava)
Area Service Manager
Meerut



LAURA

2022.09.12

08:47

DEKAPHAN LAURA

Seq.No: 0001

URINORM N

2022.09.12

09:00

.....

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	6
NIT	NEG
LEU	NEG
SG	1.025

Admin
12/9/22

QC TEST: OK
<Admin>

DEKAPHAN LAURA

Seq.No: 0002

URINORM P

2022.09.12

09:04

.....

*BLD	10	Ery/u1	1+
*BIL	3	mg/d1	2+
*UBG	1	mg/d1	1+
*KET	52	mg/d1	2+
*GLU	1000	mg/d1	3+
*PRO	100	mg/d1	
pH	9		
*NIT	POS		
*LEU	125	Leu/u1	2+
SG	1.000		

Admin
12/9/22

QC TEST: OK
<Admin>