


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 220905A/N-112	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 06 / 09 / 2022	<b>Ref. Std.</b>	: IS:12508:1988
<b>CALIBRATION DUE.</b>	: 05 / 09 / 2023	<b>C.P. No.</b>	: ACS/CP/10
<b>SRF No. &amp; DATE</b>	: 220905A/N Dt. 05.09.2022	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342722000000483F Dt. 07.09.2022	<b>Cal. At</b>	: Site / Lab

<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. C/O Chiranjeev Pathology Laboratory, Indra Chowk Hapur Road Meerut 250001
-----------------------	--

<b>ITEM CALIBRATED</b>	: Centrifuge		
Make	: Remi	Range	: (0 to 5250) RPM
L. Count	:	ID No.	: RCL/CP/Cent-01
Location	: Clinical Pathology	S. No.	: ZIBN-08838

**ENVIRONMENT CONDITION:** Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

<b>STANDARD INSTRUMENTS USED :</b>				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ TM5010EK-039	Godrej	M-210914-16-1	14.09.2022

<b>RESULTS</b>			<b>Mechanical Discipline</b>	
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)		
1000	1001.78	- 1.78		
2000	2002.14	- 2.14		
3000	3003.52	- 3.52		
4000	4005.46	- 5.46		
(average of readings)				

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

