


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 220905A/N-106	Page : 1 of 1
CALIBRATED ON. : 05 / 09 / 2022	Ref. Std. : ISO:8655:6
CALIBRATION DUE. : 04 / 03 / 2023	C.P. No. : ACS/CP/08
SRF No. & DATE : 220905A/N Dt. 05.09.2022	Phy. Inspection : O.K.
ULR No. & ISSUE DATE : CC342722000000477F Dt. 07.09.2022	Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.
C/O Chiranjeev Pathology Laboratory,
Indra Chowk Hapur Road Meerut 250001

ITEM CALIBRATED : Micro Pipette	
Make : Fine Care Corporation	Range : (100 to 1000) µl
L. Count :	ID No. : RCL/BIO/PIP-06
Location : Clinical Biochemistry	S. No. : SH88129

ENVIRONMENT CONDITION: Temp. { 25°C + 3°C } R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		100	100.39	(-) 0.39
1000		500	501.51	(-) 1.51
		1000	1003.25	(-) 3.25

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is $\pm 0.6 \mu\text{l}$ at coverage factor $k=2$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

Auth. Signatory
S.K. Sharma
CEO/TM
C-23, Sec-10, Noida