

TRICARE SERVICE SOLUTION

Puthenangady road, Kottayam : 686001.

Email Id: tricare@mail.com Ph: 0481-2934059

Mob: 8075224059

Performance Qualification

START UP

PARAMETER	Result
WBC	0.0 x 10 ³
HGB	0.0 g/dl
HCT	0.0 %
RBC	0.00 x 10 ⁶
PLT	1x10 ³

Test	QC Range Level 1	Test Results	QC Range Level 2	Test Results
WBC	6-8	7	15-17	16
HGB	10.08-13.02	12	15.5-18.5	17
RBC	5.4-6.6	6	9.9-12.01	11
PLT	210-253	245	400-447	430

CUSTOMER REPORT	INSTALLERS REPORT
	Verified all the operational qualifications. It is successful
CUSTOMER DETAILS	Hi Tech Laboratory & ECG KFPSA Shopping Complex, Near Union Bank, Kulathupalam, Kumily 685509
SIGNATURE OF CUSTOMER	
INSTALLERS DETAILS	Tricare Service Solution XV111 357/C Puthanaggady Road Kottayam 682001
SIGNATURE OF THE INSTALLER	Jobin Joseph



When you look for QUALITY IN SERVICE